
Midmark M11 - v1.0.5
Total Cycles: 1566

Name: RITTER2
Sterilizer ID: V2313228
10/05/2022 10:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.7 F 0.0
2:00 94.0 F 0.0
4:00 103.6 F 0.0
6:00 142.5 F 0.0
8:00 177.4 F 0.1
10:00 203.2 F 0.6
12:00 219.4 F 4.1
14:00 227.4 F 7.7
16:00 237.3 F 11.7
18:00 234.7 F 8.7
20:00 242.4 F 11.4
22:00 245.3 F 13.0
24:00 244.2 F 12.4
26:00 256.5 F 18.9
28:00 267.7 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.6 F 29.2
2:00 271.6 F 28.4
3:00 272.5 F 29.3
4:00 272.1 F 28.8

Min 271.4 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 207.0 F 0.0
10:00 212.8 F 0.0
15:00 214.2 F 0.0
20:00 220.4 F 0.0
25:00 225.0 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:23
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:07:26
CYCLE COMPLETE

10/05/2022 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____