
Midmark M11 - v1.0.5
Total Cycles: 3967

Name: RITTER2
Sterilizer ID: V2313228
09/23/2024 11:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 119.3 F 0.0
2:00 119.4 F 0.8
4:00 128.1 F 2.8
6:00 151.9 F 5.8
8:00 176.3 F 9.4
10:00 199.4 F 13.1
12:00 210.2 F 5.8
14:00 221.6 F 9.8
16:00 231.9 F 10.4
18:00 235.8 F 10.7
20:00 234.6 F 8.4
22:00 247.6 F 14.9
24:00 258.8 F 21.0
26:00 269.1 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 272.3 F 30.1
2:00 272.4 F 29.7
3:00 271.2 F 29.2
4:00 271.4 F 29.8

Min 271.0 F 29.1
Max 272.7 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.2 F 0.0
5:00 184.2 F 0.0
10:00 193.8 F 0.0
15:00 185.6 F 0.0
20:00 182.8 F 0.0
25:00 182.1 F 0.0
30:00 181.2 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 27:10
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:05:50
CYCLE COMPLETE

09/23/2024 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____