
Midmark M11 - v1.0.0
Total Cycles: 4312

Name: RITTER 1
Sterilizer ID: V2178198
02/20/2024 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.8 F 0.0
2:00 78.1 F 0.1
4:00 79.7 F 0.9
6:00 109.6 F 2.7
8:00 134.2 F 4.7
10:00 155.7 F 7.3
12:00 174.5 F 10.2
14:00 199.3 F 13.2
16:00 208.3 F 3.5
18:00 223.7 F 6.4
20:00 230.6 F 9.5
22:00 238.8 F 13.1
24:00 235.7 F 8.8
26:00 245.9 F 13.3
28:00 242.2 F 11.1
30:00 252.7 F 16.4
32:00 262.3 F 21.9
34:00 271.3 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.4
1:00 272.7 F 29.0
2:00 271.3 F 28.4
3:00 272.9 F 28.8
4:00 271.4 F 28.6

Min 271.2 F 28.3
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.0 F 0.0
5:00 236.2 F 0.0
10:00 199.9 F 0.0
15:00 199.0 F 0.0
20:00 196.8 F 0.0
25:00 199.9 F 0.0
30:00 196.8 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 35:19
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:13:57
CYCLE COMPLETE

02/20/2024 11:10 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____