

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4562

Name: RITTER 1  
Sterilizer ID: V2178198  
05/16/2024 10:25 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.2 F 0.0  
2:00 168.8 F 2.7  
4:00 174.6 F 6.4  
6:00 196.2 F 10.7  
8:00 216.0 F 4.5  
10:00 223.3 F 8.7  
12:00 236.4 F 12.2  
14:00 237.1 F 10.6  
16:00 235.0 F 8.5  
18:00 249.2 F 15.1  
20:00 261.1 F 21.6  
22:00 271.5 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.4  
1:00 272.7 F 29.1  
2:00 271.4 F 28.5  
3:00 272.5 F 29.4  
4:00 271.9 F 28.8

Min 271.0 F 28.3  
Max 273.0 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
5:00 210.5 F 0.0  
10:00 214.6 F 0.0  
15:00 218.3 F 0.0  
20:00 221.2 F 0.0  
25:00 227.5 F 0.0  
30:00 228.2 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 22:40  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:01:24  
CYCLE COMPLETE

05/16/2024 11:26 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_