
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002433
10:05 22/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 2:58
STERILIZING 4:02
136.0 C 221kPa 4:02
136.3 C 223kPa 4:32
136.4 C 224kPa 5:02
136.4 C 224kPa 5:32
136.4 C 223kPa 6:02
136.4 C 225kPa 6:32
136.4 C 225kPa 7:02
136.4 C 225kPa 7:32
136.4 C 225kPa 7:32
VENTING 7:32
AIR DRYING 8:27
CYCLE COMPLETE 20:22

Digital Signature #
8C91FC391C98C396
