
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.4uS / 1.5ppm
CYCLE NUMBER 003741
14:40 31/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:49
PRESSURIZING 2:55
STERILIZING 3:52
136.0 C 220kPa 3:52
136.3 C 225kPa 4:22
136.4 C 225kPa 4:52
136.3 C 224kPa 5:22
136.2 C 222kPa 5:52
136.2 C 221kPa 6:22
136.4 C 225kPa 6:52
136.0 C 221kPa 7:22
136.0 C 221kPa 7:22
VENTING 7:22
AIR DRYING 8:15
CYCLE COMPLETE 9:31

Digital Signature #
F2D5B11102D8BAAF
