
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000013
11:34 08/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:24
STERILIZING 3:20
136.0 C 220kPa 3:20
CYCLE ABORTED 3:46
STOP BUTTON PRESSED
NOT STERILE
END TIME 3:46

Digital Signature #
29DFA196150C0135
