
Midmark M11 - v1.0.0
Total Cycles: 1820

Name: RITTER 1
Sterilizer ID: V2178198
04/13/2022 12:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.1 F 0.0
2:00 166.6 F 3.0
4:00 179.3 F 7.0
6:00 212.9 F 12.5
8:00 225.7 F 6.6
10:00 236.6 F 11.5
12:00 233.5 F 8.0
14:00 243.9 F 11.7
16:00 245.0 F 12.6
18:00 257.7 F 19.4
20:00 269.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.8 F 29.1
2:00 271.8 F 28.7
3:00 272.3 F 28.5
4:00 272.6 F 29.0

Min 271.0 F 28.0
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 232.5 F 0.0
10:00 221.5 F 0.0
15:00 211.3 F 0.0
20:00 228.7 F 0.0
25:00 232.0 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 20:59
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

04/13/2022 01:29 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____