
Midmark M11 - v1.0.0
Total Cycles: 5068

Name: RITTER 1
Sterilizer ID: V2178198
11/07/2024 02:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.1 F 0.0
2:00 160.8 F 0.6
4:00 178.8 F 2.8
6:00 198.4 F 6.4
8:00 218.5 F 10.5
10:00 221.6 F 6.4
12:00 229.7 F 9.1
14:00 239.8 F 12.1
16:00 241.1 F 11.0
18:00 237.7 F 9.2
20:00 251.0 F 15.5
22:00 261.8 F 21.6
24:00 271.8 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.8 F 29.0
2:00 272.3 F 28.5
3:00 271.7 F 28.5
4:00 272.7 F 29.0

Min 271.5 F 28.2
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 254.7 F 0.0
10:00 222.9 F 0.0
15:00 209.4 F 0.0
20:00 206.3 F 0.0
25:00 222.4 F 0.0
30:00 237.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 25:28
STERILIZING: 4:00
VENTING: 3:57
DRYING: 30:00
TOTAL CYCLE: 01:04:20
CYCLE COMPLETE

11/07/2024 03:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____