
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001073
9:35 11/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:32
PRESSURIZING 3:33
STERILIZING 4:27
136.0 C 220kPa 4:27
136.2 C 223kPa 4:57
136.3 C 222kPa 5:27
136.2 C 222kPa 5:57
136.2 C 219kPa 6:27
136.4 C 225kPa 6:57
136.4 C 221kPa 7:27
136.1 C 222kPa 7:57
136.1 C 222kPa 7:57
VENTING 7:57
AIR DRYING 8:50
CYCLE COMPLETE 8:54

Digital Signature #
BF5A63F2A1757CEC
