
Midmark M11 - v1.0.5
Total Cycles: 571

Name: RITTER2
Sterilizer ID: V2313228
11/24/2021 09:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.1 F 0.0
2:00 75.6 F 0.0
4:00 87.8 F 0.0
6:00 115.3 F 0.1
8:00 150.0 F 0.2
10:00 185.4 F 0.4
12:00 202.6 F 3.8
14:00 215.3 F 7.3
16:00 225.5 F 11.1
18:00 229.3 F 7.5
20:00 237.9 F 11.5
22:00 236.4 F 9.2
24:00 240.5 F 10.2
26:00 247.3 F 14.1
28:00 258.7 F 20.2
30:00 268.7 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.8
1:00 272.5 F 29.6
2:00 271.7 F 28.7
3:00 271.9 F 29.5
4:00 272.4 F 29.2

Min 271.4 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 191.4 F 0.0
10:00 197.8 F 0.0
15:00 190.1 F 0.0
20:00 190.7 F 0.0
25:00 191.9 F 0.0
30:00 194.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 31:14
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:09:22
CYCLE COMPLETE

11/24/2021 10:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____