

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3772

Name: RITTER 1  
Sterilizer ID: V2178198  
09/14/2023 11:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
2:00 178.8 F 1.7  
4:00 191.3 F 5.4  
6:00 206.8 F 9.6  
8:00 229.4 F 7.4  
10:00 229.0 F 9.4  
12:00 238.0 F 10.2  
14:00 241.5 F 11.8  
16:00 240.2 F 10.3  
18:00 253.0 F 16.6  
20:00 264.0 F 23.0  
22:00 270.9 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.1 F 28.6  
2:00 272.5 F 29.0  
3:00 271.5 F 28.9  
4:00 272.0 F 28.7

Min 271.3 F 28.5  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.5 F 0.0  
5:00 261.6 F 0.0  
10:00 241.3 F 0.0  
15:00 227.6 F 0.0  
20:00 231.8 F 0.0  
25:00 222.7 F 0.0  
30:00 225.0 F 0.0  
Drying Complete

FILLING: 1:14  
HEATING: 22:46  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:01:42  
CYCLE COMPLETE

09/14/2023 12:29 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_