

-----  
Midmark M11 - v1.0.5  
Total Cycles: 383

Name: RITTER2  
Sterilizer ID: V2313228  
09/22/2021 10:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.2 F 0.0  
2:00 158.2 F 2.5  
4:00 171.8 F 6.9  
6:00 196.7 F 12.4  
8:00 215.1 F 7.4  
10:00 229.8 F 13.3  
12:00 237.0 F 11.3  
14:00 237.6 F 10.2  
16:00 251.7 F 17.1  
18:00 263.8 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.5  
1:00 271.9 F 29.8  
2:00 272.1 F 30.1  
3:00 271.6 F 30.0  
4:00 271.5 F 29.4  
  
Min 271.3 F 29.4  
Max 272.3 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.9 F 0.0  
5:00 223.7 F 0.0  
10:00 209.8 F 0.0  
15:00 206.4 F 0.0  
20:00 208.9 F 0.0  
25:00 205.8 F 0.0  
30:00 203.9 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 19:49  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 00:58:10  
CYCLE COMPLETE

09/22/2021 11:45 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01