

-----  
Midmark M11 - v1.0.5  
Total Cycles: 272

Name: RITTER2  
Sterilizer ID: V2313228  
08/09/2021 01:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 91.3 F 0.0  
2:00 93.2 F 0.6  
4:00 118.7 F 2.6  
6:00 147.1 F 5.3  
8:00 168.9 F 9.0  
10:00 187.4 F 13.0  
12:00 208.0 F 5.6  
14:00 217.9 F 9.4  
16:00 228.9 F 12.1  
18:00 235.6 F 10.1  
20:00 236.1 F 9.3  
22:00 245.6 F 13.6  
24:00 256.6 F 19.4  
26:00 266.6 F 25.6  
28:00 272.8 F 30.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.5  
1:00 271.9 F 29.4  
2:00 272.2 F 29.5  
3:00 272.4 F 29.7  
4:00 272.5 F 30.0

Min 271.6 F 29.4  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.9 F 0.0  
5:00 229.6 F 0.0  
10:00 220.1 F 0.0  
15:00 214.4 F 0.0  
20:00 213.9 F 0.0  
25:00 208.2 F 0.0  
30:00 207.3 F 0.0  
Drying Complete

FILLING: 1:03  
HEATING: 28:19  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:06:37  
CYCLE COMPLETE

08/09/2021 03:01 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01