

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001449  
9:06 12/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:59  
PRESSURIZING 3:01  
STERILIZING 4:02  
136.0 C 220kPa 4:02  
136.4 C 225kPa 4:32  
136.2 C 221kPa 5:02  
136.2 C 220kPa 5:32  
136.2 C 223kPa 6:02  
135.9 C 221kPa 6:32  
136.4 C 221kPa 7:02  
136.3 C 223kPa 7:32  
136.3 C 223kPa 7:32  
VENTING 7:32  
AIR DRYING 8:20  
CYCLE COMPLETE 23:21  
-----

Digital Signature #  
330EE06291782436  
-----