

-----  
Midmark M11 - v1.0.5  
Total Cycles: 71

Name: RITTER2  
Sterilizer ID: V2313228  
06/07/2021 12:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.5 F 0.0  
2:00 178.8 F 4.1  
4:00 204.4 F 10.1  
6:00 215.4 F 6.4  
8:00 232.8 F 12.4  
10:00 237.7 F 11.1  
12:00 238.2 F 10.5  
14:00 253.2 F 17.6  
16:00 265.0 F 24.8  
18:00 271.7 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.2  
1:00 271.0 F 29.0  
2:00 271.7 F 29.7  
3:00 272.7 F 30.4  
4:00 273.0 F 30.7  
  
Min 270.9 F 29.0  
Max 273.5 F 30.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
5:00 245.6 F 0.0  
10:00 224.5 F 0.0  
15:00 230.9 F 0.0  
20:00 224.6 F 0.0  
25:00 206.6 F 0.0  
30:00 191.9 F 0.0  
Drying Complete

FILLING: 2:16  
HEATING: 18:10  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:57:36  
CYCLE COMPLETE

06/07/2021 12:59 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01