
Midmark M11 - v1.0.0
Total Cycles: 3993

Name: RITTER 1
Sterilizer ID: V2178198
11/08/2023 09:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.6 F 0.2
4:00 79.1 F 1.0
6:00 104.3 F 3.0
8:00 135.1 F 5.8
10:00 163.6 F 9.2
12:00 185.4 F 13.0
14:00 202.6 F 3.3
16:00 214.5 F 6.3
18:00 223.3 F 9.7
20:00 233.5 F 12.6
22:00 235.0 F 9.3
24:00 241.4 F 11.0
26:00 243.9 F 12.1
28:00 255.4 F 18.0
30:00 265.5 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.6 F 28.9
2:00 271.8 F 28.2
3:00 271.8 F 28.6
4:00 272.4 F 28.7

Min 271.3 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.4 F 0.0
5:00 206.4 F 0.0
10:00 218.4 F 0.0
15:00 218.3 F 0.0
20:00 226.7 F 0.0
25:00 231.4 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:55
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:10:09
CYCLE COMPLETE

11/08/2023 10:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____