
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008151
9:19 05/12/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:38
STERILIZING 3:46
136.4 C 220kPa 3:46
136.3 C 223kPa 4:16
136.2 C 219kPa 4:46
136.4 C 223kPa 5:16
136.2 C 218kPa 5:46
136.1 C 220kPa 6:16
136.4 C 223kPa 6:46
136.4 C 222kPa 7:16
136.4 C 222kPa 7:16
VENTING 7:16
AIR DRYING 8:05
CYCLE COMPLETE 8:17

Digital Signature #
28A30DC19AD41A3A
