
Midmark M11 - v1.0.5
Total Cycles: 620

Name: RITTER2
Sterilizer ID: V2313228
12/09/2021 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.2 F 0.0
2:00 184.8 F 3.3
4:00 193.8 F 8.4
6:00 213.2 F 9.7
8:00 229.5 F 9.6
10:00 235.1 F 9.1
12:00 243.6 F 12.4
14:00 245.8 F 13.5
16:00 258.9 F 20.6
18:00 270.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.1 F 29.0
2:00 272.6 F 29.4
3:00 272.5 F 29.7
4:00 271.9 F 28.9

Min 271.7 F 28.8
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 215.4 F 0.0
10:00 224.8 F 0.0
15:00 226.3 F 0.0
20:00 225.2 F 0.0
25:00 224.2 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 2:50
HEATING: 18:46
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:58:40
CYCLE COMPLETE

12/09/2021 12:53 PM
ACCEPTED: X
REJECTED:
Operator: U02