
Midmark M11 - v1.0.0
Total Cycles: 4178

Name: RITTER 1
Sterilizer ID: V2178198
01/04/2024 09:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 70.8 F 0.2
4:00 80.6 F 1.0
6:00 110.1 F 2.9
8:00 140.4 F 5.3
10:00 164.7 F 8.3
12:00 183.4 F 11.7
14:00 208.2 F 2.2
16:00 216.9 F 5.4
18:00 226.2 F 8.4
20:00 234.8 F 11.9
22:00 230.8 F 6.9
24:00 241.4 F 11.6
26:00 236.5 F 8.7
28:00 248.8 F 14.4
30:00 259.2 F 20.1
32:00 268.9 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.6 F 28.8
2:00 272.1 F 29.0
3:00 272.5 F 28.6
4:00 272.5 F 29.2

Min 271.3 F 28.3
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.2 F 0.0
5:00 238.2 F 0.0
10:00 227.0 F 0.0
15:00 232.4 F 0.0
20:00 232.1 F 0.0
25:00 227.3 F 0.0
30:00 210.6 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 33:54
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:12:16
CYCLE COMPLETE

01/04/2024 10:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____