
Midmark M11 - v1.0.5
Total Cycles: 290

Name: RITTER2
Sterilizer ID: V2313228
08/19/2021 10:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.5 F 0.0
2:00 184.9 F 4.2
4:00 205.3 F 10.0
6:00 221.5 F 6.0
8:00 236.0 F 12.0
10:00 238.4 F 10.7
12:00 238.7 F 10.3
14:00 253.8 F 17.7
16:00 266.1 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.0
1:00 272.0 F 28.9
2:00 272.4 F 29.2
3:00 272.8 F 29.6
4:00 272.1 F 29.5

Min 271.8 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 199.3 F 0.0
10:00 201.2 F 0.0
15:00 199.8 F 0.0
20:00 195.6 F 0.0
25:00 192.2 F 0.0
30:00 189.3 F 0.0
Drying Complete

FILLING: 3:23
HEATING: 17:28
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

08/19/2021 11:41 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____