
Midmark M11 - v1.0.0
Total Cycles: 5301

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2025 03:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.8 F 0.0
2:00 141.3 F 3.1
4:00 159.0 F 6.7
6:00 185.3 F 10.7
8:00 202.0 F 4.0
10:00 217.2 F 8.1
12:00 230.5 F 12.9
14:00 233.1 F 9.1
16:00 238.3 F 10.7
18:00 243.2 F 12.5
20:00 254.8 F 18.5
22:00 265.4 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.1 F 29.6
2:00 272.7 F 29.8
3:00 272.8 F 29.6
4:00 272.1 F 29.3

Min 271.3 F 28.9
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.8 F 0.0
5:00 204.9 F 0.0
10:00 199.1 F 0.0
15:00 191.8 F 0.0
20:00 187.8 F 0.0
25:00 189.9 F 0.0
30:00 198.0 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 23:59
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:02:45
CYCLE COMPLETE

01/30/2025 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____