

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1695

Name: RITTER 1  
Sterilizer ID: V2178198  
03/02/2022 03:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.5 F 0.0  
2:00 197.6 F 3.6  
4:00 196.8 F 8.4  
6:00 226.2 F 9.0  
8:00 230.7 F 9.4  
10:00 234.6 F 8.8  
12:00 243.3 F 12.7  
14:00 241.8 F 11.3  
16:00 255.1 F 18.1  
18:00 266.8 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.4 F 29.1  
2:00 271.8 F 28.4  
3:00 271.4 F 28.6  
4:00 272.5 F 29.1  
  
Min 271.1 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.3 F 0.0  
5:00 234.6 F 0.0  
10:00 232.5 F 0.0  
15:00 223.5 F 0.0  
20:00 218.2 F 0.0  
25:00 225.5 F 0.0  
30:00 221.7 F 0.0  
Drying Complete

FILLING: 1:58  
HEATING: 19:35  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:42  
CYCLE COMPLETE

03/02/2022 04:34 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_