
Midmark M11 - v1.0.0
Total Cycles: 1546

Name: RITTER 1
Sterilizer ID: V2178198
01/20/2022 12:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.9 F 0.0
2:00 192.3 F 4.2
4:00 197.2 F 8.9
6:00 227.6 F 8.9
8:00 231.9 F 9.6
10:00 231.0 F 8.5
12:00 243.6 F 12.8
14:00 242.2 F 11.3
16:00 255.4 F 18.1
18:00 267.1 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.5 F 29.4
2:00 272.0 F 28.7
3:00 272.6 F 29.1
4:00 272.4 F 29.4

Min 270.9 F 28.1
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.0 F 0.0
5:00 235.4 F 0.0
10:00 231.4 F 0.0
15:00 228.9 F 0.0
20:00 231.3 F 0.0
25:00 230.0 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

01/20/2022 01:24 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____