
Midmark M11 - v1.0.0
Total Cycles: 2140

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2022 03:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.4 F 0.0
2:00 184.1 F 3.8
4:00 196.5 F 8.5
6:00 225.8 F 10.3
8:00 232.0 F 8.3
10:00 242.0 F 13.5
12:00 240.0 F 10.6
14:00 235.5 F 8.4
16:00 250.0 F 15.3
18:00 262.0 F 21.9
20:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.8 F 29.0
2:00 271.6 F 28.1
3:00 272.4 F 29.0
4:00 272.3 F 28.7

Min 271.2 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 236.9 F 0.0
10:00 228.7 F 0.0
15:00 234.7 F 0.0
20:00 223.3 F 0.0
25:00 224.1 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:38
CYCLE COMPLETE

07/15/2022 04:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____