
Midmark M11 - v1.0.5
Total Cycles: 3430

Name: RITTER2
Sterilizer ID: V2313228
03/04/2024 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 68.9 F 0.0
4:00 73.7 F 0.0
6:00 97.1 F 0.8
8:00 136.7 F 3.2
10:00 164.8 F 6.3
12:00 183.6 F 9.7
14:00 200.5 F 11.3
16:00 215.8 F 6.0
18:00 232.1 F 9.7
20:00 236.9 F 10.4
22:00 237.5 F 10.7
24:00 234.3 F 8.1
26:00 248.0 F 14.5
28:00 259.4 F 20.7
30:00 269.5 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.7 F 29.5
2:00 271.6 F 29.4
3:00 272.9 F 29.4
4:00 271.6 F 29.5

Min 271.3 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 172.8 F 0.0
5:00 231.0 F 0.0
10:00 202.5 F 0.0
15:00 176.4 F 0.0
20:00 171.4 F 0.0
25:00 168.5 F 0.0
30:00 166.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:51
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:56
CYCLE COMPLETE

03/04/2024 10:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____