
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 008505
10:53 03/04/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:22
STERILIZING 3:15
136.4 C 221kPa 3:15
136.4 C 224kPa 3:45
136.1 C 224kPa 4:15
136.4 C 221kPa 4:45
136.4 C 223kPa 5:15
136.4 C 224kPa 5:45
136.4 C 218kPa 6:15
136.4 C 223kPa 6:45
136.4 C 223kPa 6:45
VENTING 6:45
AIR DRYING 7:32
CYCLE COMPLETE 37:32

Digital Signature #
408FB58506B1A5EF
