

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3646

Name: RITTER2  
Sterilizer ID: V2313228  
05/27/2024 09:19 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.2 F 0.0  
2:00 70.6 F 0.0  
4:00 74.2 F 0.1  
6:00 91.0 F 1.5  
8:00 130.2 F 3.7  
10:00 158.8 F 6.0  
12:00 177.3 F 8.5  
14:00 201.1 F 11.5  
16:00 207.7 F 3.8  
18:00 212.1 F 6.8  
20:00 222.2 F 10.3  
22:00 230.2 F 9.3  
24:00 235.9 F 10.6  
26:00 235.6 F 8.9  
28:00 245.7 F 13.9  
30:00 255.9 F 19.3  
32:00 265.3 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 271.3 F 29.5  
2:00 273.0 F 30.2  
3:00 271.6 F 29.6  
4:00 272.7 F 30.0  
  
Min 271.0 F 29.2  
Max 273.1 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.0 F 0.0  
5:00 202.1 F 0.0  
10:00 190.7 F 0.0  
15:00 182.0 F 0.0  
20:00 180.7 F 0.0  
25:00 180.4 F 0.0  
30:00 188.9 F 0.0  
  
Drying Complete

FILLING: 0:49  
HEATING: 33:57  
STERILIZING: 4:00  
VENTING: 3:41  
DRYING: 30:00  
TOTAL CYCLE: 01:12:28  
CYCLE COMPLETE

05/27/2024 10:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_