
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002398
8:52 20/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:02
STERILIZING 3:55
136.0 C 220kPa 3:55
136.4 C 225kPa 4:25
136.4 C 224kPa 4:55
136.3 C 223kPa 5:25
136.3 C 223kPa 5:55
136.3 C 220kPa 6:25
136.0 C 220kPa 6:55
136.2 C 223kPa 7:25
136.2 C 223kPa 7:25
VENTING 7:26
AIR DRYING 8:14
CYCLE COMPLETE 23:58

Digital Signature #
0D24DB2B97B80F56
