

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3845

Name: RITTER 1  
Sterilizer ID: V2178198  
10/02/2023 08:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.8 F 0.0  
2:00 88.9 F 0.0  
4:00 82.1 F 0.0  
6:00 98.3 F 0.0  
8:00 133.2 F 1.4  
10:00 162.0 F 3.1  
12:00 186.8 F 5.4  
14:00 216.9 F 9.1  
16:00 219.0 F 12.8  
18:00 223.9 F 5.8  
20:00 234.9 F 9.7  
22:00 237.7 F 10.6  
24:00 240.7 F 10.8  
26:00 235.9 F 8.4  
28:00 249.9 F 14.9  
30:00 260.8 F 21.0  
32:00 270.7 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 27.9  
1:00 272.5 F 28.6  
2:00 271.3 F 27.9  
3:00 272.4 F 28.4  
4:00 271.5 F 28.1  
  
Min 271.3 F 27.7  
Max 272.6 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.2 F 0.0  
5:00 215.3 F 0.0  
10:00 230.4 F 0.0  
15:00 218.5 F 0.0  
20:00 229.3 F 0.0  
25:00 227.0 F 0.0  
30:00 227.5 F 0.0  
  
Drying Complete

FILLING: 0:47  
HEATING: 32:38  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:10:45  
CYCLE COMPLETE

10/02/2023 09:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_