
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001528
10:21 10/25/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:07
PRESSURIZING 6:58
STERILIZING 8:36
136.0 C 220kPa 8:36
136.0 C 219kPa 9:06
136.0 C 218kPa 9:36
136.2 C 220kPa 10:06
136.3 C 219kPa 10:36
136.0 C 217kPa 11:07
136.4 C 222kPa 11:37
136.5 C 220kPa 12:06
VENTING 12:06
AIR DRYING 13:11
CYCLE COMPLETE 43:11

Digital Signature #
7A9836FC3C7547F5
