

-----  
Midmark M11 - v1.0.5  
Total Cycles: 543

Name: RITTER2  
Sterilizer ID: V2313228  
11/11/2021 11:46 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 169.2 F 0.0  
2:00 176.5 F 3.8  
4:00 199.4 F 9.0  
6:00 216.8 F 7.4  
8:00 234.3 F 10.4  
10:00 233.1 F 7.8  
12:00 240.4 F 10.6  
14:00 247.7 F 14.3  
16:00 260.4 F 21.4  
18:00 271.9 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 272.4 F 29.1  
2:00 272.0 F 29.4  
3:00 271.7 F 28.9  
4:00 272.0 F 29.5  
  
Min 271.2 F 28.6  
Max 272.6 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
5:00 233.6 F 0.0  
10:00 200.3 F 0.0  
15:00 195.4 F 0.0  
20:00 192.0 F 0.0  
25:00 189.8 F 0.0  
30:00 189.7 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 18:30  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:40  
CYCLE COMPLETE

11/11/2021 12:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_