
Midmark M11 - v1.0.5
Total Cycles: 189

Name: RITTER2
Sterilizer ID: V2313228
07/12/2021 06:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.7 F 0.0
2:00 74.4 F 0.0
4:00 81.4 F 0.8
6:00 95.1 F 2.7
8:00 124.3 F 5.2
10:00 153.2 F 8.3
12:00 184.4 F 11.7
14:00 210.7 F 3.9
16:00 212.2 F 7.6
18:00 227.6 F 11.4
20:00 225.6 F 7.2
22:00 236.0 F 12.3
24:00 238.3 F 10.4
26:00 250.1 F 16.2
28:00 260.2 F 22.1
30:00 270.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 30.0
1:00 272.0 F 29.9
2:00 272.0 F 29.9
3:00 272.4 F 30.0
4:00 272.4 F 30.2

Min 271.7 F 29.9
Max 272.8 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 203.9 F 0.0
10:00 209.9 F 0.0
15:00 189.9 F 0.0
20:00 207.6 F 0.0
25:00 234.8 F 0.0
30:00 231.8 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 31:33
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:09:57
CYCLE COMPLETE

07/12/2021 07:35 AM
ACCEPTED: X
REJECTED:
Operator: U01