
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000909
15:41 30/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:35
STERILIZING 4:34
136.0 C 220kPa 4:34
136.1 C 221kPa 5:04
136.4 C 225kPa 5:34
136.4 C 225kPa 6:04
136.4 C 223kPa 6:34
136.2 C 222kPa 7:04
136.3 C 223kPa 7:34
136.1 C 223kPa 8:04
136.1 C 223kPa 8:04
VENTING 8:04
AIR DRYING 8:57
CYCLE COMPLETE 9:01

Digital Signature #
6C1113E699AAF0DC
