

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1599

Name: RITTER 1  
Sterilizer ID: V2178198  
02/07/2022 10:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.5 F 0.0  
2:00 171.7 F 3.4  
4:00 183.3 F 7.7  
6:00 207.2 F 12.8  
8:00 224.4 F 7.0  
10:00 237.8 F 12.2  
12:00 235.5 F 9.3  
14:00 240.7 F 9.5  
16:00 247.8 F 14.1  
18:00 260.0 F 20.8  
20:00 271.0 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.5 F 28.9  
2:00 271.5 F 28.2  
3:00 272.6 F 29.0  
4:00 272.0 F 28.4  
  
Min 271.4 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.2 F 0.0  
5:00 232.0 F 0.0  
10:00 226.3 F 0.0  
15:00 229.6 F 0.0  
20:00 226.8 F 0.0  
25:00 231.5 F 0.0  
30:00 228.6 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 20:44  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

02/07/2022 11:36 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_