
Midmark M11 - v1.0.0
Total Cycles: 3226

Name: RITTER 1
Sterilizer ID: V2178198
05/12/2023 01:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.2 F 0.0
2:00 195.1 F 0.0
4:00 206.7 F 4.6
6:00 226.1 F 9.6
8:00 231.5 F 7.7
10:00 237.3 F 11.0
12:00 235.8 F 8.4
14:00 240.9 F 10.3
16:00 247.6 F 14.0
18:00 260.2 F 20.8
20:00 271.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.3 F 28.6
2:00 271.7 F 28.6
3:00 272.1 F 28.5
4:00 271.5 F 28.4

Min 271.4 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 231.0 F 0.0
10:00 231.0 F 0.0
15:00 229.2 F 0.0
20:00 230.0 F 0.0
25:00 227.8 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

05/12/2023 02:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____