
Midmark M11 - v1.0.5
Total Cycles: 3823

Name: RITTER2
Sterilizer ID: V2313228
07/31/2024 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.0 F 0.0
2:00 78.5 F 0.0
4:00 90.7 F 0.0
6:00 113.2 F 0.2
8:00 149.6 F 0.8
10:00 178.4 F 3.9
12:00 213.8 F 7.2
14:00 217.6 F 10.9
16:00 220.1 F 6.1
18:00 229.6 F 8.2
20:00 239.9 F 12.3
22:00 235.3 F 9.1
24:00 240.2 F 10.6
26:00 246.0 F 13.5
28:00 257.7 F 19.7
30:00 268.2 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.5 F 29.0
2:00 272.2 F 29.4
3:00 271.8 F 28.6
4:00 272.5 F 29.4

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 230.7 F 0.0
10:00 224.1 F 0.0
15:00 208.4 F 0.0
20:00 204.0 F 0.0
25:00 206.0 F 0.0
30:00 199.4 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:16
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:09:29
CYCLE COMPLETE

07/31/2024 10:07 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____