
Midmark M11 - v1.0.5
Total Cycles: 2621

Name: RITTER2
Sterilizer ID: V2313228
07/07/2023 09:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.3 F 0.0
2:00 71.3 F 0.0
4:00 95.4 F 0.0
6:00 121.6 F 0.0
8:00 168.7 F 0.4
10:00 190.4 F 0.8
12:00 205.0 F 4.6
14:00 216.4 F 8.1
16:00 227.9 F 12.1
18:00 229.5 F 7.7
20:00 240.6 F 12.4
22:00 238.7 F 10.2
24:00 236.6 F 8.5
26:00 250.5 F 15.7
28:00 262.1 F 22.4
30:00 272.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 271.7 F 28.6
2:00 271.7 F 28.6
3:00 271.8 F 28.5
4:00 272.0 F 28.6

Min 271.7 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 233.3 F 0.0
5:00 208.7 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 30:16
STERILIZING: 4:00
VENTING: 2:44
DRYING: 5:04
TOTAL CYCLE: 00:43:12

07/07/2023 10:25 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____