

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3656

Name: RITTER2  
Sterilizer ID: V2313228  
05/29/2024 11:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 122.7 F 0.0  
2:00 122.3 F 0.3  
4:00 124.4 F 2.2  
6:00 149.5 F 5.0  
8:00 189.6 F 8.5  
10:00 197.4 F 12.6  
12:00 211.8 F 5.7  
14:00 220.8 F 9.8  
16:00 231.7 F 9.9  
18:00 235.4 F 10.7  
20:00 237.0 F 8.3  
22:00 246.4 F 14.4  
24:00 257.2 F 20.2  
26:00 267.2 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.7  
1:00 272.7 F 30.0  
2:00 271.8 F 30.1  
3:00 272.3 F 29.8  
4:00 272.2 F 30.3

Min 271.4 F 29.6  
Max 272.9 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 212.3 F 0.0  
10:00 211.5 F 0.0  
15:00 205.0 F 0.0  
20:00 210.8 F 0.0  
25:00 224.5 F 0.0  
30:00 228.3 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 27:35  
STERILIZING: 4:00  
VENTING: 3:39  
DRYING: 30:00  
TOTAL CYCLE: 01:06:07  
CYCLE COMPLETE

05/29/2024 12:09 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_