
Midmark M11 - v1.0.0
Total Cycles: 1610

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2022 03:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.8 F 0.0
2:00 169.1 F 3.3
4:00 178.0 F 7.4
6:00 207.7 F 12.7
8:00 220.6 F 6.5
10:00 234.7 F 11.4
12:00 233.0 F 8.1
14:00 242.9 F 11.6
16:00 244.6 F 12.7
18:00 257.1 F 19.3
20:00 268.2 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.4 F 29.0
2:00 272.0 F 28.9
3:00 272.4 F 28.9
4:00 271.7 F 28.7

Min 271.6 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 230.4 F 0.0
10:00 225.6 F 0.0
15:00 226.7 F 0.0
20:00 230.4 F 0.0
25:00 228.4 F 0.0
30:00 224.7 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 21:13
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:54
CYCLE COMPLETE

02/08/2022 04:38 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____