
Midmark M11 - v1.0.5
Total Cycles: 818

Name: RITTER2
Sterilizer ID: V2313228
02/22/2022 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 81.0 F 0.0
4:00 114.5 F 0.0
6:00 146.6 F 0.1
8:00 183.4 F 0.3
10:00 206.0 F 0.6
12:00 216.3 F 1.6

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 219F
CHAMBER PRESS 2.3PSI
** ITEMS NOT STERILE **

02/22/2022 09:56 AM