
Midmark M11 - v1.0.0
Total Cycles: 5333

Name: RITTER 1
Sterilizer ID: V2178198
02/11/2025 12:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.3 F 0.0
2:00 106.8 F 0.7
4:00 117.3 F 2.8
6:00 147.2 F 5.2
8:00 168.8 F 7.9
10:00 186.2 F 11.1
12:00 216.0 F 4.7
14:00 210.8 F 5.7
16:00 224.4 F 9.5
18:00 237.2 F 12.4
20:00 235.5 F 9.5
22:00 241.2 F 10.5
24:00 244.3 F 12.2
26:00 254.7 F 17.7
28:00 264.3 F 23.4
30:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 271.6 F 28.8
2:00 272.7 F 29.4
3:00 272.7 F 29.2
4:00 271.9 F 28.8

Min 270.9 F 28.2
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 198.1 F 0.0
10:00 210.0 F 0.0
15:00 218.6 F 0.0
20:00 204.7 F 0.0
25:00 201.3 F 0.0
30:00 202.0 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 30:26
STERILIZING: 4:00
VENTING: 3:59
DRYING: 30:00
TOTAL CYCLE: 01:09:10
CYCLE COMPLETE

02/11/2025 01:41 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____