
Midmark M11 - v1.0.5
Total Cycles: 1701

Name: RITTER2
Sterilizer ID: V2313228
11/08/2022 04:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.5 F 0.0
2:00 177.7 F 0.4
4:00 202.0 F 3.9
6:00 220.2 F 9.1
8:00 234.0 F 9.0
10:00 237.2 F 11.5
12:00 238.0 F 9.5
14:00 238.1 F 9.1
16:00 250.9 F 15.6
18:00 263.1 F 22.7
20:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.3 F 28.9
2:00 271.6 F 28.6
3:00 272.4 F 29.0
4:00 271.7 F 28.4

Min 271.5 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.7 F 0.0
5:00 221.1 F 0.0
10:00 232.8 F 0.0
15:00 229.8 F 0.0
20:00 225.7 F 0.0
25:00 221.1 F 0.0
30:00 216.9 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:08
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

11/08/2022 05:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____