
Midmark M11 - v1.0.5
Total Cycles: 2778

Name: RITTER2
Sterilizer ID: V2313228
08/16/2023 01:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.8 F 0.0
2:00 134.3 F 0.0
4:00 149.4 F 0.0
6:00 178.4 F 0.8
8:00 198.5 F 4.5
10:00 212.2 F 8.6
12:00 226.0 F 13.2
14:00 234.3 F 9.3
16:00 237.7 F 10.2
18:00 243.1 F 13.1
20:00 243.9 F 12.4
22:00 256.0 F 18.9
24:00 266.8 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.5 F 29.5
2:00 271.4 F 28.7
3:00 272.0 F 29.6
4:00 271.9 F 29.1

Min 271.0 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.5 F 0.0
5:00 237.3 F 0.0
10:00 223.7 F 0.0
15:00 196.1 F 0.0
20:00 180.8 F 0.0
25:00 177.1 F 0.0
30:00 174.9 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 25:37
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:03:52
CYCLE COMPLETE

08/16/2023 02:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____