

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3939

Name: RITTER 1  
Sterilizer ID: V2178198  
10/23/2023 05:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	118.9 F	0.0
2:00	121.4 F	0.3
4:00	148.0 F	3.3
6:00	175.1 F	7.0
8:00	197.8 F	11.7
10:00	212.7 F	4.6
12:00	223.2 F	9.3
14:00	234.8 F	9.9
16:00	238.8 F	10.9
18:00	236.2 F	8.7
20:00	250.3 F	15.4
22:00	262.1 F	22.1
24:00	272.6 F	28.9

STERILIZING

mm:ss	Degrees	PSI
0:00	271.5 F	28.3
1:00	272.4 F	28.8
2:00	271.5 F	28.4
3:00	272.5 F	28.9
4:00	271.6 F	28.2

Min 271.4 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING

Drying Start	mm:ss	Degrees	PSI
	0:00	196.8 F	0.0
	5:00	214.6 F	0.0
	10:00	229.7 F	0.0
	15:00	227.7 F	0.0
	20:00	227.1 F	0.0
	25:00	228.7 F	0.0
Drying Complete	30:00	227.0 F	0.0

FILLING: 1:17  
HEATING: 24:19  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:02:59  
CYCLE COMPLETE

10/23/2023 06:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_