

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4561

Name: RITTER 1  
Sterilizer ID: V2178198  
05/16/2024 09:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 173.7 F 0.0  
2:00 171.2 F 2.3  
4:00 172.6 F 5.8  
6:00 191.4 F 9.8  
8:00 210.4 F 8.2  
10:00 218.5 F 7.8  
12:00 229.6 F 12.2  
14:00 232.7 F 8.6  
16:00 242.5 F 11.9  
18:00 243.7 F 12.3  
20:00 255.4 F 18.3  
22:00 265.8 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.5 F 29.3  
2:00 271.9 F 28.7  
3:00 271.4 F 28.7  
4:00 272.6 F 29.3

Min 271.2 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.5 F 0.0  
5:00 222.2 F 0.0  
10:00 228.2 F 0.0  
15:00 232.7 F 0.0  
20:00 227.3 F 0.0  
25:00 227.9 F 0.0  
30:00 228.9 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 23:50  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:02:55  
CYCLE COMPLETE

05/16/2024 10:04 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_