
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001123
9:47 20/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:18
STERILIZING 4:31
136.0 C 220kPa 4:31
136.3 C 225kPa 5:01
136.4 C 225kPa 5:31
136.3 C 224kPa 6:01
136.4 C 226kPa 6:31
136.3 C 223kPa 7:01
136.4 C 226kPa 7:31
136.3 C 224kPa 8:01
136.3 C 224kPa 8:01
VENTING 8:01
AIR DRYING 8:55
CYCLE COMPLETE 9:00

Digital Signature #
8F2047898CB11342
