
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004322
11:50 08/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:25
STERILIZING 3:26
136.0 C 220kPa 3:26
136.4 C 226kPa 3:56
136.4 C 225kPa 4:26
136.4 C 225kPa 4:56
136.4 C 224kPa 5:26
136.4 C 224kPa 5:56
136.4 C 223kPa 6:26
136.4 C 225kPa 6:56
136.4 C 225kPa 6:56
VENTING 6:56
AIR DRYING 8:16
CYCLE COMPLETE 8:29

Digital Signature #
95437254D81B475F
