
Midmark M11 - v1.0.5
Total Cycles: 3500

Name: RITTER2
Sterilizer ID: V2313228
04/05/2024 11:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.5 F 0.0
2:00 155.7 F 2.1
4:00 172.8 F 5.4
6:00 194.0 F 9.2
8:00 227.4 F 13.5
10:00 225.7 F 7.3
12:00 235.1 F 11.8
14:00 233.9 F 8.8
16:00 240.4 F 11.1
18:00 245.5 F 13.5
20:00 257.0 F 19.7
22:00 267.4 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.5 F 29.8
2:00 271.9 F 29.2
3:00 271.4 F 29.4
4:00 272.4 F 29.8

Min 271.2 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.2 F 0.0
5:00 231.2 F 0.0
10:00 216.5 F 0.0
15:00 214.6 F 0.0
20:00 190.0 F 0.0
25:00 183.6 F 0.0
30:00 181.4 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 23:27
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:02:25
CYCLE COMPLETE

04/05/2024 12:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____