
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001596
17:34 15/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:48
STERILIZING 3:50
136.3 C 224kPa 3:50
136.3 C 225kPa 4:20
136.4 C 226kPa 4:50
136.3 C 224kPa 5:20
136.3 C 225kPa 5:50
136.2 C 225kPa 6:20
136.4 C 226kPa 6:50
136.2 C 224kPa 7:20
136.2 C 224kPa 7:20
VENTING 7:20
AIR DRYING 8:11
CYCLE COMPLETE 8:18

Digital Signature #
E91E831AD361A122
